



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

REQUEST FOR QUOTATION

Date: 10/21/2022

Quotation No:2022-0228

Company Name: _____

Address: _____

Business Permit No.: _____

TIN: _____

PhilGEPS Registration No.(required): _____

The City Government of Muntinlupa, through its Bids and Awards Committee, intends to procure for the **PURCHASE OF ELECTROCARDIOGRAM WITH DATA VIEWER TO BE USED BY OSMUN** which will be undertaken in accordance with **Section 53.9** of the 2016 Revised Implementing Rules and Regulations of Republic Act No.9184.

Please quote your **best offer** for the item/s described herein, subject to the Terms and Conditions provided. Submit your quotation duly signed by you or your duly authorized representative not later than the **DEADLINE** on **October 27, 2022**.

A copy of the following documents are also required to be submitted along with your quotation/proposal:

1. Mayor's/Business Permit: (Certified True Copy)	3. PhilGEPS Registration (Certified True Copy)
2. Accomplished and Notarized Omnibus Sworn Statement (Original)	4. Certificate of Registration (Certified True Copy)

Quotations/Proposals must be submitted to the BAC Office of the City Government of Muntinlupa for checking & validation.

For any clarification, you may contact **Bids & Awards Committee** at telephone no.(02)8861-1127



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

INSTRUCTIONS:

- (2) Do not alter the contents of this in any way.
- (3) technical specifications with asterisks(*) are mandatory. Failure to comply with any of the mandatory requirements will disqualify your
- (4) Failure to follow these instructions will disqualify your entire quotation.

After having carefully read and accepted the Terms and Conditions, I/we submit our quotation/s for the item/s as follows:

Procurement Project	Approved Budget for the Contract (ABC)
PURCHASE OF ELECTROCARDIOGRAM WITH DATA VIEWER TO BE USED BY OSMUN	Two Hundred Thousand Pesos (Php200,000.00)

Technical Specifications:

QTY	UNIT OF ISSUE	ITEM DESCRIPTION	Compliance		REMARKS
			Yes	No	
1	unit	ELECTROCARDIOGRAM WITH DATAVIEWER			
		Specifications:			
		a. Input Impedence : >10 MΩ			
		b. Leads : Standard 12 leads and Cabrera lead			
		c. Simultaneous or Real time input			
		d. Patient Input : One patient			
		e. Common Mode rejection : > 100dB			
		f. Frequency Response : 0.05 – 150Hz			
		g. (Within -3dB +0.5dB)			
		h. Time Constant : > 3.2 seconds			
		i. Digital Filter : AC : -30dB or less, 60, 50Hz			
		j. Drift : -3dB, 0.5Hz			
		k. Muscle : -3dB, 35Hz, 25Hz			
		l. A/D Conversion : 13bit			
		m. Sampling Rate : 1000/sec and Pacemaker of 100μs			
		n. Defibrillator Protection : To 5000V, 400J			
		o. Patient Leakage Current : 10μA or less			
		p. Display : Touch Screen 5.7 inches STN LCD with Monochrome 320x240 dots and ECG Display of 4/12-Lead			
		q. Print : Printer: Thermal Array 8 dots/mm. Sensitivity: 2.5, 5, 10 or 20 mm/mV. Auto Selection Speed: 5, 10, 12.5, 25 or 50 mm/sec. Recording Paper: R80x30R-C (Roll)			
		r. Internal Memory: 200 resting ECG			



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

		s. Data Entry : Patient information entry by touching alphanumeric screen			
		t. Operation mode : Auto, Freeze, Pre-check, Post Stress, Manual, Lead - off Alarm. Arrhythmia monitoring mode : 30 sec, 1-5 min, VPC/SVPC summary, Autonomic nervous system test			
		u. Language : English			
		v. Safety : Class I Type CF			
		w. Power : Power of AC 110 – 120V, 220 – 240V, 50 – 60Hz, a Nickel Hydride battery that operates continuously for 2 hours. To be charged for hours.			
		x. Size : Dimension of W266 x D173x H75 mm that weights approximately 1.5 kg w/ battery			
		y. 5.7 inch touch screen for patient data entry and preview of 12 lead ECG and analysis results with Huge Data Capability. Data transmission with LAN port XML data format output is possible. Direct connection with Epson Inkjet Printer. USB port for connecting with external keyboard.			
		II. INCLUSIONS:			
1	unit	Power Cable			
1	unit	AC adaptor			
1	unit	Limb Electrode			
1	unit	Chest Electrode			
1	unit	Patient Cable			
1	unit	70g ECG electrode cream			
1	unit	Operation Manual			
1	unit	CD-R			
1	unit	ECG recording paper (R80 x 30R – C)			
1	unit	Paper Spool			
1	unit	ECG cart			
		Delivery and installation must be done within Thirty (30) calendar days			
		Comprehensive 12 months warranty			



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

		2 years of free labor			
		Availability of spare parts within 5 – 7 years after purchase			
		With back-up unit in case of malfunction / breakdown			
		With Operation and User Manual in English.			
		With manufacturer and Distributors Warranty.			
		With Schedule of preventive maintenance			
		With Preventive Maintenance Certification and Calibration Sticker during maintenance of unit.			
		Supplier / Distributor must have at least five (5) years of Operation in the Philippines.			



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

Your Quotation						
Summary of Approved Budget			OFFERED QUOTATION			
QTY	UNIT OF ISSUE	ITEM DESCRIPTION	QTY	UNIT OF ISSUE	ITEM DESCRIPTION	TOTAL PRICE
1	unit	ELECTROCARDIOGRAM WITH DATAVIEWER				
		Specifications:				
		a. Input Impedence : >10 MΩ				
		b. Leads : Standard 12 leads and Cabrera lead				
		c. Simultaneous or Real time input				
		d. Patient Input : One patient				
		e. Common Mode rejection : > 100dB				
		f. Frequency Response : 0.05 – 150Hz				
		g. (Within -3dB +0.5dB)				
		h. Time Constant : > 3.2 seconds				
		i. Digital Filter : AC : -30dB or less, 60, 50Hz				
		j. Drift : -3dB, 0.5Hz				
		k. Muscle : -3dB, 35Hz, 25Hz				
		l. A/D Conversion : 13bit				
		m. Sampling Rate : 1000/sec and Pacemaker of 100μs				
		n. Defibrillator Protection : To 5000V, 400J				
		o. Patient Leakage Current : 10μA or less				
		p. Display : Touch Screen 5.7 inches STN LCD with Monochrome 320x240 dots and ECG Display of 4/12-Lead				
		q. Print : Printer: Thermal Array 8 dots/mm. Sensitivity: 2.5, 5, 10 or 20 mm/mV. Auto Selection Speed: 5, 10, 12.5, 25 or 50 mm/sec. Recording Paper: R80x30R-C (Roll)				



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

		r. Internal Memory: 200 resting ECG				
		s. Data Entry : Patient information entry by touching alphanumeric screen				
		t. Operation mode : Auto, Freeze, Pre-check, Post Stress, Manual, Lead - off Alarm. Arrhythmia monitoring mode : 30 sec, 1-5 min, VPC/SVPC summary, Autonomic nervous system test				
		u. Language : English				
		v. Safety : Class I Type CF				
		w. Power : Power of AC 110 – 120V, 220 – 240V, 50 – 60Hz, a Nickel Hydride battery that operates continuously for 2 hours. To be charged for hours.				
		x. Size : Dimension of W266 x D173x H75 mm that weights approximately 1.5 kg w/ battery				
		y. 5.7 inch touch screen for patient data entry and preview of 12 lead ECG and analysis results with Huge Data Capability. Data transmission with LAN port XML data format output is possible. Direct connection with Epson Inkjet Printer. USB port for connecting with external keyboard.				
		II. INCLUSIONS:				
1	unit	Power Cable				
1	unit	AC adaptor				
1	unit	Limb Electrode				
1	unit	Chest Electrode				
1	unit	Patient Cable				
1	unit	70g ECG electrode cream				
1	unit	Operation Manual				
1	unit	CD-R				



Republic of the Philippines
City Government of Muntinlupa

National Road Putatan Muntinlupa City

BIDS and AWARDS COMMITTEE

www.muntinlupacity.gov.ph

1	unit	ECG recording paper (R80 x 30R – C) Roll				
1	unit	Paper Spool				
1	unit	ECG cart				
		Delivery and installation must be done within Thirty (30) calendar days				
		Comprehensive 12 months warranty				
		2 years of free labor				
		Availability of spare parts within 5 – 7 years after purchase				
		With back-up unit in case of malfunction / breakdown				
		With Operation and User Manual in English.				
		With manufacturer and Distributors Warranty.				
		With Schedule of preventive maintenance				
		With Preventive Maintenance Certification and Calibration Sticker during maintenance of unit.				
		Supplier / Distributor must have at least five (5) years of Operation in the Philippines.				
		TOTAL	200,000.00	Total Offered quotation (in Php)	PhP	



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

TERMS AND CONDITIONS:

1. Bidders shall provide correct and accurate information required in this form.
2. Bidders must quote for all or all the items.
3. Price quotation/s must be valid for a period of thirty (30) calendar days from the date of submission.
4. Price quotation/s, to be denominated in Philippine peso shall include all taxes, duties and/or levies payable.
5. Quotations exceeding the Approved Budget for the contract shall be rejected.
6. Award of contract shall be made to the lowest quotation (for goods and infrastructure) or, the highest rated offer (for consulting services) which complies with the minimum technical specifications and other terms and conditions stated herein.
7. Any interlineations, erasures overwriting shall be valid only if they are signed or initialed by you or any of your duly authorized representative/s.
8. The Item/s shall be delivered according to the requirements specified in the Technical Specifications.
9. The GSO shall have the right to inspect and/or to test the goods to confirm their conformity to the technical specifications.

Signature over Printed Name

Position/Designation

Office Telephone No.

Mobile Phone No./Fax No.

Email address/es