



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

REQUEST FOR QUOTATION

Date: 04/18/2023
Quotation No:2023-0154

Company Name: _____

Address: _____

Business Permit No.: _____

TIN: _____

PhilGEPS Registration No.(required): _____

The City Government of Muntinlupa, through its Bids and Awards Committee, intends to procure **Purchase of Various Equipment to be used on Water Laboratory of CHO**, which will be undertaken in accordance with **Section 53.9** of the 2016 Revised Implementing Rules and Regulations of Republic Act No.9184.

Please quote your **best offer** for the item/s described herein, subject to the Terms and Conditions provided.

A copy of the following documents are also required to be submitted along with your quotation/proposal:

1. Mayor's/Business Permit: (Certified True Copy)	3. PhilGEPS Registration (Certified True Copy)
2. Accomplished and Notarized Omnibus Sworn Statement (Original)	4. Certificate of Registration (Certified True Copy)

Quotations/Proposals must be submitted to the BAC Office of the City Government of Muntinlupa for checking & validation.

For any clarification, you may contact **Bids & Awards Committee** at telephone no.(02)8861-1127

INSTRUCTIONS:

- (2) Do not alter the contents of this in any way.
- (3) technical specifications with asterisks(*) are mandatory. Failure to comply with any of the mandatory requirements will disqualify your
- (4) Failure to follow these instructions will disqualify your entire quotation.

After having carefully read and accepted the Terms and Conditions, I/we submit our quotation/s for the item/s as follows:

Procurement Project	Approved Budget for the Contract (ABC)
Purchase of Various Equipment to be used on Water Laboratory of CHO	One Hundred Forty Thousand Pesos

Technical Specifications:

QTY	UNIT OF ISSUE	ITEM DESCRIPTION	Compliance		REMARKS
			Yes	No	
1	unit	Stirring Hotplate			
		Plate Size: 10 x 10"			
		Top Material: Ceramic, square			
		Max Container size: 6L			
		Stir Range: 50-1500RPM			



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

		(or more)			
		increments			
		accuracy: Temperature accuracy within			
		±2°C and stirring speeds at ±5%			
		(41.1 x 28.7 x 10.2 cm)			
		Display: Digital display			
		automatic over-temperature protection			
		Certifications/Compliance: CE			
		Magnetic Stir Bar			
		<i>Stir Bar retriever</i>			
		and services			
		Micropipette, 20-200 uL			
		Type: Single Channel			
		Volume Range: 20-200 uL			
		Increments: 0.2 uL			
		Volume Display: 4-digit display			
		Autoclavable at 121°C,			
		integrated calibration function,			
		ejector			
		with pipette tips, quality			
		Certificate, and operating manual			
		and services			
1	unit	Laboratory Refrigerator, Upright			
		Temperature range: 2-8°C			
		Temperature accuracy: 0.1°C			
		Capacity: 238-260 L			
		Shelves: at least 4			
		Door: Glass door with electric heating			
		large screen LED display			
		and low temperature, Power			
		failure alarm, door ajar			
		Alarm: Thermostat failure			
		refrigeration system			
		Refrigerant: R600a, CFC Free			
		Defrost: No frost or auto defrost			
		Inclusion: Compatible AVR			
		services			
1	unit	Unstirred Water Bath with Cover			



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

		Capacity: 5.5 L			
		Temperature range : RT+5-99°C			
		Temperature stability: ±0.3°C or better			
		Temp error alarm: ±2°C			
		stainless steel			
		Display: LCD screen			
		Timing range: 1-9999min			
		Internal Dimensions: 130*280*150mm			
		Power consumption: 500 W			
		Features/s: Water bath cover			
		Overtemperature protection			
		Warranty: at least 1 year on parts and services			
Summary of Approved Budget			OFFERED QUOTATION		
QTY	UNIT OF ISSUE	ITEM DESCRIPTION	QTY	UNIT OF ISSUE	TOTAL PRICE
1	unit	Stirring Hotplate			
		Plate Size: 10 x 10"			
		Top Material: Ceramic, square			
		Max Container size: 6L			
		Stir Range: 50-1500RPM			
		Temperature Range: ambient to 400°C (or more)			
		Temperature Adjustment: 1 degree increments			
		Temperature and stirring speed accuracy: Temperature accuracy within ±2°C and stirring speeds at ±5%			
		Dimensions (LxWxH): 16.2 x 11.3 x 4 in (41.1 x 28.7 x 10.2 cm)			
		Display: Digital display			
		Features: HOT TOP warning display, automatic over-temperature protection			
		Certifications/Compliance: CE			
		Inclusions: 1" and 2" PTFE-Coated Magnetic Stir Bar			
		Stir Bar retriever			
		Warranty: at least one (1) year on parts and services			
		Micropipette, 20-200 uL			
		Type: Single Channel			



Republic of the Philippines
City Government of Muntinlupa
 National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

		Volume Range: 20-200 uL			
		Increments: 0.2 uL			
		Volume Display: 4-digit display			
		Pipette tip compatibility: Universal, Autoclavable at 121°C,			
		Volume-change protection, integrated calibration function,			
		Corrosion-resistant piston and ejector			
		Inclusions: silicone grease, sample bag with pipette tips, quality			
		Certificate, and operating manual			
		Warranty: at least one (1) year on parts and services			
1	unit	Laboratory Refrigerator, Upright			
		Temperature range: 2-8°C			
		Temperature accuracy: 0.1°C			
		Capacity: 238-260 L			
		Shelves: at least 4			
		Door: Glass door with electric heating			
		Control System: Microprocessor control, large screen LED display			
		Audible and visual alarm for: High and low temperature, Power			
		failure alarm, door ajar			
		Alarm: Thermostat failure			
		Refrigerator type: Forced air refrigeration system			
		Refrigerant: R600a, CFC Free			
		Defrost: No frost or auto defrost			
		Inclusion: Compatible AVR			
		Warranty: at least 1 year on parts and services			
1	unit	Unstirred Water Bath with Cover			
		Capacity: 5.5 L			
		Temperature range : RT+5-99°C			
		Temperature stability: ±0.3°C or better			
		Temp error alarm: ±2°C			
		Chamber material: corrosion-resistant stainless steel			
		Display: LCD screen			



Republic of the Philippines
City Government of Muntinlupa
National Road Putatan Muntinlupa City
BIDS and AWARDS COMMITTEE
www.muntinlupacity.gov.ph

		Timing range: 1-9999min				
		Internal Dimensions: 130*280*150mm				
		Power consumption: 500 W				
		Features/s: Water bath cover				
		Overttemperature protection				
		Warranty: at least 1 year on parts and services				
			140,000.00	Total Offered quotation (in Php)	Php	

TERMS AND CONDITIONS:

1. Bidders shall provide correct and accurate information required in this form.
2. Bidders must quote for all or all the items.
3. Price quotation/s must be valid for a period of thirty (30) calendar days from the date of submission.
4. Price quotation/s, to be denominated in Philippine peso shall include all taxes, duties and/or levies payable.
5. Quotations exceeding the Approved Budget for the contract shall be rejected.
6. Award of contract shall be made to the lowest quotation (for goods and infrastructure) or, the highest rated offer (for consulting services) which complies with the minimum technical specifications and other terms and conditions stated herein.
7. Any interlineations, erasures overwriting shall be valid only if they are signed or initialed by you or any of your duly authorized representative/s.
8. The Item/s shall be delivered according to the requirements specified in the Technical Specifications.
9. The GSO/Engineering Office shall have the right to inspect and/or to test the goods to confirm their conformity to the technical specifications.

Signature over Printed Name

Position/Designation

Office Telephone No.

Mobile Phone No./Fax No.

Email address/es